

# Certificate of Compliance

**Certificate:** 1099249

**Master Contract:** 162509 (LR 90209)

**Project:** 1099249

**Date Issued:** March 22, 2001

**Issued to:** Theralase Inc.  
20 Ferrier St., Unit 1  
Markham, ON L3R 3Z5  
Canada  
**Attention:**


*The products listed below are eligible to bear the CSA Mark shown,  
with adjacent indicator "C" and "US".*



**Issued by:**

  
P. Cheung, P.Eng.

**Authorized by:**

  
Nick Alfano,  
Operations Manager

## PRODUCTS

CLASS 8750 01 - MEDICAL ELECTRICAL EQUIPMENT

CLASS 8750 81 - MEDICAL ELECTRICAL EQUIPMENT - CERTIFIED TO U.S. STANDARDS

Therapeutic Medical Laser, Model TLC-1000, rated 100-240V, 6W, 50-60Hz, used with power supply Model TLC-154 (rated input 100-240V, 50-60Hz, 1.2A, output +16V dc, 2.81A), Continuous Operation, Class II, Applied Part Type BF, Equipment Protection IPX0.

The "C" and "US" indicator adjacent to the CSA Mark signifies that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S. This includes products eligible to bear the NRTL indicator. NRTL, i.e. Nationally Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognised to perform certification to U.S. Standards.